



Dissolution Accessories,
PO Box 173,
4900 AD,
Oosterhout,
The Netherlands

Tel + 31 (0) 162-471485
Fax + 31 (0) 162-471486

CERTIFICATE OF ANALYSIS

PRODUCT: POTASSIUM PHOSPHATE pH 7.20 @25°C
DISSOLUTION MEDIA CONCENTRATE

PRODUCT No.: DMC26-25

LOT NO: DMC2614D1

DATE OF TEST: 01/05/2014

EXPIRY DATE: 28/04/2016

MEAN pH VALUE: pH 7.180 @25°C

SPECIFICATION: pH 7.18 - 7.22 @25°C after dilution

DENSITY (g/ml): 1.1527 @25°C

COMPOSITION: Potassium dihydrogenorthophosphate
Sodium Hydroxide

pH MEASUREMENT: Tested in accordance with in-house method TPPHB.
Measured with a combination glass electrode after multiple
point calibration with reference materials.

TRACEABILITY: Standard buffer solutions are certified traceable to the
following U.S. National Institute of Standards and
Technology, SRM 185h Potassium Hydrogen Phthalate,
SRM 186-I-g Potassium Dihydrogen Phosphate and SRM
186-II-g Disodium Hydrogen Phosphate.

This certificate relates solely to the lot number given above. The uncertainty of measurement has been calculated not to exceed ± 0.02 pH units at 95% confidence level.

This product complies with the USP requirements for dissolution media when diluted with purified water as instructed.

Approved By: Quality Manager

Darren Mc Jate

Date: 20/06/2014

Supplement to Certificate issued on 02/05/2014