

CERTIFICATE OF ANALYSIS

PRODUCT: SODIUM LAURYL SULPHATE 0.50%
DISSOLUTION MEDIA CONCENTRATE

PRODUCT No.: DMC029-10

LOT NO: DMC29121B1

DATE OF TEST: 25/02/2021

EXPIRY DATE: 28/02/2023

MEAN NORMALITY: 0.0173N, after dilution.

SPECIFICATION: 0.0170 – 0.0177N, after dilution.

METHOD: Standardised in accordance with specifications ASTM
E200-08.

REFERENCE: Standard Method of Preparation, Standardisation and
Storage of Standard Solutions for Chemical Analysis.
ASTM E200-08

COMPOSITION: Sodium Lauryl Sulphate
Purified Water

This product complies with the USP & CP requirements for dissolution media when diluted with purified water as instructed.

Approved By:



QC Technician

Date: 25/02/2021