

Dissolution Accessories,
PO Box 173,
4900 AD,
Oosterhout,
The Netherlands

Tel + 31 (0) 162-471485
Fax + 31 (0) 162-471486

CERTIFICATE OF ANALYSIS

PRODUCT: SODIUM PHOSPHATE pH 6.80 @25°C
DISSOLUTION MEDIA CONCENTRATE

PRODUCT No.: DMC034-25

LOT NO: DMC3416D1

DATE OF TEST: 28th April 2016

EXPIRY DATE: 28th April 2018

MEAN pH VALUE: 6.82 @25°C

SPECIFICATION: pH 6.78 - 6.82 @25°C after dilution

DENSITY (g/ml): 1.1205g/ml @25°C

COMPOSITION: Sodium Dihydrogen Phosphate
Sodium Hydroxide

pH MEASUREMENT: Tested in accordance with in-house method TPPHB.
Measured with a combination glass electrode after multiple
point calibration with reference materials.

TRACEABILITY: Standard buffer solutions are certified traceable to the
following U.S. National Institute of Standards and
Technology, SRM 185i Potassium Hydrogen Phthalate,
SRM 186-I-g Potassium Dihydrogen Phosphate and SRM
186-II-g Disodium Hydrogen Phosphate.

This certificate relates solely to the lot number given above. The uncertainty of measurement has been calculated not to exceed ± 0.02 pH units at 95% confidence level.

This product complies with the USP requirements for dissolution media when diluted with purified water as instructed.

Approved By: QC Technician
Date: 28th April 2016

Derdre MacLigan