

Certificate of Analysis

PRODUCT: Phosphate pH 6.5 @ 25°C (FASSIF, no enzyme)
Dissolution Media Concentrate

PRODUCT No.: DMF560-5L

LOT NO.: FASF15C1

DATE OF TEST: 08-03-2018

EXPIRT DAYE: 28-03-2019

MEAN pH-value: 6.6@25°C

pH Measurement: The result reported above was determined by analysis of a sample of this lot taken at time of manufacture. Test Method used was TPPHG. Measured with a combination glass electrode after multiple point calibration with reference materials. It is certified traceable to the following National Institute of Standards and Technology (USA), SRM 185h Potassium Hydrogen Phthalate, SRM 186-I-g Potassium Dihydrogen Phosphate and SRM 186-II-g Disodium Hydrogen Phosphate.

Approved by: QC Technician

Date: 08-03-2018

This certificate relates solely to the lot number given above. The uncertainty of measurement has been calculated not to exceed ± 0.02 pH units at 95% confidence level, i.e. coverage factor $k = 2$.

Dissolution accessories is a ProSense BV Brand

ProSense BV • P.O. Box 173 • 4900 AD Oosterhout • The Netherlands
www.dissolutionaccessories.com • info@dissolutionaccessories.com